

3/27 (金)



電圧降下測定 (Comparator at threshold)

125V
125V

MRPC

測定値

Com	LV [V]	LV [V]	LV [A]
	0.5 (0.499)	0.51	0.00 0.01
	1.0 (1.002)	1.02	0.01
	1.5 (1.498)	1.53	0.02
	2.0 (1.999)	2.05	0.03
	2.5 (2.501)	2.56	0.037
	3.0 (3.002)	3.08	0.04



MRPC

設定-測定値

LVの表示

Comparator [V]	設定	LV [V]	LV [A]	目盛 (設定値)
0.503	0.518	0.52	0.01	
1.000	1.0432	1.04	0.02	
1.500	1.5514	1.56	0.03	
2.003	2.0790	2.09	0.04	
2.504	2.5989	2.62	0.05	
3.000	3.1173	3.14	0.06	

TSC (Comparator 3)

Comparator [V]	設定	LV [V]	LV [A]	目盛 (設定値)
0.994	1.0218	1.02	0.01	
1.999	2.0633	2.06	0.02	
3.009	3.1103	3.10	0.03	
4.004	4.1426	4.13	0.04	
5.004	5.1789	5.17	0.05	
6.001	6.214	6.20	0.06	
7.000	7.2417	7.23	0.07 0.07	
7.99	8.278	8.26	0.08	
8.99	9.312	9.30	0.09	
10.01	10.370	10.35	0.11	

ESC (Comparator 4)

Comparator [V]	LV [V]		LV [A]
	FLA- 目盛	目盛	目盛 (2桁)
1.001	1.0374	0.4	0
2.003	2.0827	1.4	0
3.000	3.1229	2.4	0
4.002	4.1694	3.5	0
5.001	5.2119	4.5	0
6.003	6.257	5.6	0
7.00	7.295	6.6	0
8.00	8.339	7.6	0
9.00	9.382	8.7	0
10.01	10.438	9.7	0

