Symposium on Nuclear Data 2020



Contribution ID: 34

Type: Poster Presentation

Study on characteristics of neutron and γ-ray fields at compact neutron source RANS-II facility by simulation by the PHITS code / PHITS による p-Li 反 応を用いた中性子源 RANS-II の放射線場特性に関する 研究

Thursday, 26 November 2020 17:09 (1h 41m)

RIKEN Accelerator driven compact Neutron Source-II (RANS-II) based on the ⁷Li(p, n)⁷Be reaction for neutron production with 2.49 MeV proton beam, has been under beam commissioning to demonstrate specific performance of the system. RIKEN has a prospect of realizing novel non-destructive neutron inspection for infrastructures with the use of RANS. As prominent characteristics, RANS-II has the maximum neutron energy of 0.8 MeV, which is lower than that of 5 MeV at RANS based on the ⁹Be(p, n)⁹B reaction with 7 MeV proton injection, and gives extremely forward favored angular distribution with respect to the proton beam direction. Also, it should be emphasized that RANS-II system is installed in a relatively small space isolated by concrete shield with boron containment. Accordingly, there should be quite large differences in neutronic performances between RANS-II and RANS in terms of neutron spectrum and angular distributions. In preparation of experiments at RANS-II, the simulation of radiation fields for neutron and γ -ray in RANS-II experimental hall plays a critical important role for designing experimental set-up in low background.

Then, we have performed simulations to characterize radiation fields of RANS-II The cross section libraries implemented in PHITS are utilized in neutron and γ -ray transportations. Several important conditions of RANS-II modeling are as follows:

• The lithium (Li) target is made by depositing thin Li layer of about 100 μ m on a 5 mm thick Cu substrate cooled by water in the target station.

• The target station with about 90 cm side cubic shape, configures five layers; polyethylene, lead, borated polyethylene, lead and iron, to reduce the radiation leakage.

• There is a hole with a 15 \times 15 $\rm cm^2$ cross section in the forward direction.

• The experimental hall has dimensions of $14 \times 5.5 \times 3.0 \text{ m}^3$ surrounded by floor and wall made of concrete (partly the borated concrete) and polyethylene ceiling.

As a result, the neutron and γ -ray distribution spreads widely in the experimental hall due to the wide openings in the target station. The scattered radiation could be the major contributor to the background of experiments. On the other hand, primary neutrons produced at the target are shielded reasonably by the target station. To design effective collimators for high quality beam extraction, we have calculated neutron beam profiles with parameters of collimator diameters and materials. It is shown that there is an optimized collimator configuration to extract suitable beam.

Primary author: Mr SUGIHARA / 杉原, Kenta / 健太 (RIKEN Nishina Center, RIKEN, 2-1 Hirosawa Wako, Saitama 351-0198, Japan / 理化学研究所仁科加速器科学研究センター, Department of Applied Quantum Physics and Nuclear Engineering, Kyushu University, Motooka, Nishi-ku, Fukuoka 819-0395, Japan / 九州大学大学院工 学研究院エネルギー量子工学専攻)

Co-authors: Dr IKEDA / 池田, Yujiro / 裕二郎 (RIKEN Center for Advanced Photonics, RIKEN, 2-1 Hirosawa,

Wako, Saitama 351-0198, Japan / 理化学研究所光量子工学研究センター中性子ビーム技術開発チーム, J-PARC Center, Japan Atomic Energy Agency, 2-4 Shirakata, Tokai-mura, Naka-gun, Ibaraki 319-1195, Japan / 日本 原子力研究開発機構 J-PARC センター); Dr KOBAYASHI / 小林, Tomohiro / 知洋 (RIKEN Center for Advanced Photonics, RIKEN, 2-1 Hirosawa, Wako, Saitama 351-0198, Japan / 理化学研究所光量子工学研究センター中性 子ビーム技術開発チーム); Dr FUJITA / 藤田, Kunihiro / 訓裕 (RIKEN Center for Advanced Photonics, RIKEN, 2-1 Hirosawa, Wako, Saitama 351-0198, Japan / 理化学研究所光量子工学研究センター中性子ビーム技術開発 チーム); Dr SHIGYO / 執行, Nobuhiro / 信寛 (Department of Applied Quantum Physics and Nuclear Engineering, Kyushu University, Motooka, Nishi-ku, Fukuoka 819-0395, Japan / 九州大学大学院工学研究所仁科加速器科学研究 センター); Dr TANAKA / 田中, Kanenobu / 鐘信 (RIKEN Nishina Center / 理化学研究所仁科加速器科学研究 センター); Dr OTAKE / 大竹, Yoshie / 淑恵 (RIKEN Center for Advanced Photonics, RIKEN, 2-1 Hirosawa, Wako, Saitama 351-0198, Japan / 理化学研究所光量子工学研究センター中性子ビーム技術開発

Presenter: Mr SUGIHARA / 杉原, Kenta / 健太 (RIKEN Nishina Center, RIKEN, 2-1 Hirosawa Wako, Saitama 351-0198, Japan / 理化学研究所仁科加速器科学研究センター, Department of Applied Quantum Physics and Nuclear Engineering, Kyushu University, Motooka, Nishi-ku, Fukuoka 819-0395, Japan / 九州大学大学院工学研究院エネルギー量子工学専攻)

Session Classification: Poster