

## Development of a neutron detector for nuclear data measurement using high-intensity neutron beam

Thursday, 26 November 2020 17:00 (1h 47m)

Highly precise neutron nuclear data is required in nuclear transmutation research of long-lived minor actinides (MA) in nuclear waste. In neutron capture cross section measurement, monitoring the number of the incident neutrons is necessary. However, in measurement with J-PARC/ANNRI, direct neutron monitoring system has not been employed. To make measurement with ANNRI more robust, an additional neutron beam monitor is required. Conventional neutron detectors cannot be used as a beam monitor at ANNRI because of two reasons, high counting rate environment and gamma-flash. The neutron flux at ANNRI is one of the highest in the world. Gamma-flash, an intense gamma-ray burst produced when the proton beam pulse bombards the spallation target, can paralyze a detector generally used in nuclear data measurement. In general, a semiconductor detector or an inorganic scintillator, which is adopted for a neutron detector, has relatively longer response time and is unsuitable for beam monitoring at ANNRI.

Therefore, a combination of a thin plastic scintillator and a  $^6\text{LiF}$  foil was selected as a detection system, whose fast response enabled detecting neutrons at a high counting rate. Low gamma ray sensitivity of a thin plastic scintillator allows measuring fast TOF region without count loss or detector paralysis. The geometry of the  $^6\text{LiF}$  foil, the plastic scintillator, and photomultiplier tube (PMT) was designed. The optimal thickness of the  $^6\text{LiF}$  foil was determined with simulation codes, SRIM and PHITS. A  $^6\text{LiF}$  foil was made by vacuum deposition method. A test detector system was built to study the feasibility of the method.

The detector system was tested under the high neutron irradiation condition at J-PARC /ANNRI. A neutron TOF spectrum was successfully measured without significant count loss or detector paralysis. A neutron energy spectrum was driven from difference of TOF spectrum with and without  $^6\text{LiF}$ . The neutron spectrum was compared with a past neutron spectrum and good agreement was obtained. Statistic error was 0.68 % at 6.0 meV even though measurement times in this study was pretty short (~11 min).

**Primary author:** NAKANO, Hideto (Tokyo Institute of Technology )

**Co-authors:** Prof. KATABUCHI, Tatsuya (Tokyo Institute of Technology); ROVIRA LEVERONI, Gerard (JAEA); KODAMA, Yu (Tokyo Tech); TERADA, Kazushi (Kyoto University); Dr KIMURA, Atsushi (JAEA); Dr NAKAMURA, Shoji (JAEA); Dr ENDO, Shunsuke (JAEA)

**Presenter:** NAKANO, Hideto (Tokyo Institute of Technology )

**Session Classification:** Poster